

CFEZ Cover Glass Not Attached

Unfortunately a user cannot reattach the glass window that is used to seal the optics from the measuring environment. Please contact Service@HunterLab.com to request an SRO. Include your contact information and instrument serial number. If you would like to rent an instrument to use while yours is being repaired please make the request in the e-mail.