

Why doesn't my sample start and stop at the same location after a measurement?

The turntable begins rotation before the sensor begins taking measurements to ensure that the sample is moving at the desired uniform speed. A complete rotation will occur from the time the sensor begins measurements.